

Role of Novel Gnathodynamometer for Measurement of Bite Force in Adult Population of North India

Dr. Munna Khan, Dr. Syed Ansar Ahmad, Dr. Mohammad Faisal
Corresponding Author: Dr. Syed Ansar Ahmad

Aim; The aim of the present study is to measure the maximum bite force in the natural dentition with the help of bite force measurement device.

Materials and methods; 100 subjects randomly selected from the population aged between 18-60 years were tested. The device was placed in between upper and lower central incisors and patients were asked to bite until the first sign of pain was felt. This exercise was done twice and mean was taken. Same procedure was followed for right and left first molar.

Results; The data proved that maximum bite force was more in young males than in young females. Also bite force tends to decrease as age progresses.

Conclusion; Bite force is more in young males than in young females. Also bite force on the right molars was similar to left molars.

Date of Submission: 20-05-2021

Date of Acceptance: 05-06-2021

I. Introduction;

Materials and methods; 100 subjects were randomly selected by minimisation technique from the Out Patient Department of Jamia Millia Islamia, New Delhi. Bite measurement device was prepared for the purpose of the study consisting of two forks and sensor in between.

Each patient was informed in advance about the purpose of the study. It was explained that the biting on the force measuring device should continue only until first signs of pain are felt. Initial measurement was done on central incisor region than on left and right molars.

Introduction; A major indicator of the functional state of the masticatory system is the generated biting force.¹ The ability to bite is a function of the craniomandibular structures, including the muscles of mastication, the temporomandibular joint and the dentition, be it natural or artificial. The magnitude of masticatory force is an indicator of functional status of the masticatory system. The effect of chewing efficiency and bite force on oral health-related quality was evaluated in the field of geriatric and implant dentistry [1,2]. Measuring the masticatory force of individuals has been widely used to understand the biomechanical principles of masticatory muscles and outcomes of prosthodontics treatments. Also, the masticatory force is very important for diagnosis and treatment of dysfunction and behavior of the stomatognathic system [3].

Method and objectives of the study were thoroughly explained to all participants and written informed consent was obtained. Age, gender, the dominant hand, facial height and Angle's class of occlusion were recorded. With the patient in a seated position and occlusal plane parallel to the horizontal plane, the sensor was placed in the patient's mouth in-between the following teeth in an orderly fashion: 1. Occlusal surfaces of the maxillary right first molar and mandibular right first molar; 2. Occlusal surfaces of the maxillary left first molar and mandibular left first molar; 3. Incisal edges of the maxillary and mandibular anterior teeth. The patients were asked to press their teeth in maximum intercuspation. By doing so, the maximum masticatory force in right and left molars, and also in the anterior segment was measured and recorded. Series of three recordings were taken. The rest period of one minute was given between each recording to prevent muscle fatigue. After every 10 measurements, the functions of the device and sensors were tested using a mass with a specific weight.



Fig;1.Gnatho dynamometer for bite force record



Fig;2 bite force being measured

II. Results

This study was performed on 50 males and 50 females. The females had a mean age of 19-55 years .The males were in the range of 19-55 years. The mean force measured at the right side was 22.4 kg in males and 16.2 kg in females .

Maximum bite force in males (mean of 50 patients)

Central incisors 8.6 kg

First molar (right)20.4

First molar (left) 21.2

Maximum bite force in females(mean of 50 patients)

Central incisors 8.2 kgs

First molar(right) 18.1kgs

First molar (left)16.1 kgs

III. Discussion

Among the 50 male patients on which study was conducted in the department ,bite force in males was 8.6 kgs and for females it was 8.2kgs ,which was almost equal .For right first molar (females)MBF was 18.1kgs and for right first molar (males)MBF was 20.4 . in males for first molar (left side) MBF was 21.2 and

for females first molar(left)MBF was 16.1kgs.Findings suggest that bite force on the central incisor for both males and females was similar ,whereas for molars ,bite force was higher for males than for females. Also bite force tend to increase in patients from 20-35 and it reduced after reaching 40years. Several clinical and experimental studies have shown the significant role played by masticatory muscle function in the craniofacial growth.

In 1681 ,Borelli was among the first to measure the force of mastication using a somewhat similar technique of assessing the bite force by transducer placed between one pair of opposing teeth ,leaving the rest of the dentition separated.

Strain gauges were used by Howell and Manly, Floystrand et al., and Bakke et al. for measuring bite forces.^[8] The measurement of bite forces has been remained a matter of interest among many researchers. However, there is inconsistency in the findings and maximum value of bite forces presented by different authors.^[7,8] The reasons of this variation may be many such as the device used to record the bite force, its sensitivity, comfort of the volunteer, and psychological state of volunteer. In addition, genetic and ethnic, food habits, and geographical factors may be also responsible for this variation. Individual neuromuscular mechanism may itself be also an important factor for this difference

The most successful of the entire lot of bite force recorders consisted of metallic face ,an electronic instrument ,and instant standardization device .This forms the basis of the bite force recorder. To reduce the metallic impact on teeth and to prevent cross contamination ,disposable caps for fork were used.

Previous studies have evaluated the effects of several factors such as craniofacial morphology ^[10-11], gender weight and height and pattern of occlusal contacts on masticatory loads. Gender differences are the most effective on masticatory loads. According to most researchers, masticatory forces are higher in males as compared to females. This difference is probably due to the difference in muscle strength or size of teeth in males and females ^[12]. However, in the study by Abu Alhaija et al, in 2010 no difference was reported in the masticatory forces between males and females. The obtained results revealed that males had a significantly higher masticatory force than females, which confirms previous findings in this regard ^[13]. However, Abu Alhaija et al, in 2010 found no significant difference in maximum masticatory forces of men and women. It appears that the effect of gender on the difference in the masticatory forces is attributed to the higher muscular strength in men ^[15]; which per se is due to anatomical differences between the two sexes ^[15].

Ferrario et al^[1], in most masticatory forces were measured to be in the range of 446N to 1200N with very high accuracy and low (3%) error rate. The maximum values are 228N in men and 198N in women. According to M. Manoj Srikanth, B. Pavan Kumar, V. Venkatesh, V. Vidya Devi Evaluation of Bite Forces in Healthy Individuals Indian J Dent Adv 2018; 10(3): 101-104.The maximum bite force was measured in young healthy volunteers and it was found that the average bite forces in incisor region were 10.6 kg, right molar region 38.5 kg, and left molar region were 40.1 kg^[16]. In a study of incisal biting forces using a strain gauge gnathodynamometer on a group of young males and females between the ages of 10 and 25 years, Garner and Kotwal^[17] reported a mean biting force of 25–01 lb (11-3 kg) ± 14–11 lb (6–4 kg). They found that males bite harder than females and that, in agreement with Worner and Anderson, biting forces increase with age up to adolescence.[12] In a similar study, Gupta et al. found that voluntary bite force in a healthy adult was on the order of 15.4 Kp in the incisor and 48.3 and 49.2 Kp in the left and right molar regions, respectively. Kshirsagar et al^[18] revealed that the maximum voluntary bite force measurement in healthy Indian individuals is of the order of 36 kg in the molar region and 15 kg in the incisor region.

Conflict of interest: None

References;

- [1]. Fernandes CP, Glantz PO, Svensson SA, Bergmark A. A novel sensor for bite force determinations. Dent Mater. 2003;19(2):118–126
- [2]. Ingervall B, Minder C. Correlation between maximum bite force and facial morphology in children. Angle Orthod 1997;67:415-422.
- [3]. Kogawa EM, Calderon PS, Lauris JR, et al. Evaluation of maximal bite force in temporomandibular disorders patients. J Oral Rehabil 2006;33:559-565.
- [4]. Mantyvaara J, Sjöholm T, Kirjavainen T, et al. Altered control of submaximal bite force during bruxism in humans. Eur J Appl Physiol Occup Physiol 1999;79:325-330.
- [5]. Nickolay Apostolov, Ivan Chakalov, Todor Drajev. Measurement of the Maximum Bite Force in the Natura. Ferrario VF, Sforza C, Serrao G, Dellavia C, Tartaglia GM. Single tooth bite forces in healthy young adults. J Oral Rehabil 2004;31:18-22.
- [6]. Bakke M, Holm B, Jensen BL, Michler L, Möller E. Unilateral, isometric bite force in 8-68-year-old women and men related to occlusal factors. Scand J Dent Res 1990;98:149-58.
- [7]. Dean JS, Throckmorton GS, Ellis E 3rd, Sinn DP. A preliminary study of maximum voluntary bite force and jaw muscle efficiency in pre-orthognathic surgery patients. J Oral Maxillofac Surg 1992;50:1284-8.
- [8]. Howell AH, Manly RS. An electronic strain gauge for measuring oral forces. J Dent Res 1948;27:705-12. 6. Fløystrand F, Kleven E, Oilo G. A novel miniature bite force recorder and its clinical application. Acta Odontol Scand 1982;40:209-14.
- [9]. Manoj Srikanth, B. Pavan Kumar, V. Venkatesh, V. Vidya Devi Evaluation of Bite Forces in Healthy Individuals Indian J Dent Adv 2018; 10(3): 101-104.The

- [10]. Kshirsagar R, Jaggi N, Halli R. Bite force measurement in mandibular parasymphyseal fractures: A preliminary clinical study. *Craniomaxillofac Trauma Reconstr* 2011;4:241-41 Dentition with a Gnathodynamometer I S S U E 2 , 2 0 1 4. MedInform.
- [11]. Koc D, Dogan A, Bek B. Bite force and influential factors on bite force measurements: A literature review. *Eur J Dent* 2010;4:223-32.
- [12]. Ferrario VF, Sforza C, Zanotti G, et al. Maximal bite forces in healthy young adults as predicted by surface electromyography. *J Dent* 2004;32:451-457.
- [13]. Floystrand F, Kleven E, Oilo G. A novel miniature bite force recorder and its clinical application. *Acta Odontol Scand* 1982;40:209-214.
- [14]. Fogle LL, Glaros AG. Contributions of facial morphology, age, and gender to EMG activity under biting and resting conditions: a canonical correlation analysis. *J Dent Res* 1995;74:1496-1500
- [15]. *J Clin Exp Dent*. 2018 Nov; 10(11): e1063–e1068. Elham Abu-Alhaija, 1 Arwa I. Owais, 2 and Hiba Obaid 3. Maximum occlusal bite force in pre-school children with different occlusal patterns
- [16]. 9. Manoj Srikanth, B. Pavan Kumar, V. Venkatesh, V. Vidya Devi Evaluation of Bite Forces in Healthy Individuals *Indian J Dent Adv* 2018; 10(3): 101-104.
- [17]. L D Garner, N S Kotwal . *J Dent Res* Jul-Aug 1973;52(4):698-702. doi: 10.1177/00220345730520041001. Correlation study of incisive biting forces with age, sex, and anterior occlusion.
- [18]. Kshirsagar R, Jaggi N, Halli R. Bite force measurement in mandibular parasymphyseal fractures: A preliminary clinical study. *Craniomaxillofac Trauma Reconstr* 2011;4:241-41 Dentition with a Gnathodynamometer I S S U E 2 , 2 0 1 4. MedInform

Dr. Munna Khan, et. al. "Role of Novel Gnathodynamometer for Measurement of Bite Force in Adult Population of North India." *IOSR Journal of Dental and Medical Sciences (IOSR-JDMS)*, 20(06), 2021, pp. 01-04.